

**Substances that elicit positive and false-positive results using  
Siemens EMIT® II Plus reagents**

**Emit® II Plus Propoxyphene Assay (300ng cutoff)** *Calibrated with Propoxyphene at 300 ng/ml*

| <b>Drug</b>     | <b>ng/ml</b> |
|-----------------|--------------|
| Propoxyphene    | 300          |
| Norpropoxyphene | 800          |

**Known False Positives:** None