

**Substances that elicit positive and false-positive results using  
Siemens EMIT® II Plus reagents**

**Emit® II Plus Cocaine Metabolite Assay (150 ng cutoff)** *Calibrated with Benzoylecgonine at 150 ng/ml*

<b>Drug</b>	<b>ng/ml</b>
Benzoylecgonine	150
Cocaine	18,000-53,000
Ecgonine	2,000-6,000

**False Positives: None**